


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMR-1-P-EA-EW-ED-TRE-ECL-M1-TM*

S.N.: *5222*

Date: *0 LT 30 89*



*40 db ATTEN*  
*.5 VOLTS/DIV*  
*7.0 MHz*  
*1.0V*

- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*5 TO 50 NS*
- c) Rise time:  
*≤ 100 PSEZ ON*
- d) Fall time:  
*≤ 1 NS*
- e) PRF:  
*≤ 135 PSEZ ON*  
*0 TO 10 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.000*  
*300 mA (MAX)*