


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-pw-c-f*
S.N.: *5223*
Date: *SEPT 14 1989*



R_L = 50 Ω
PRF ≈ 200 Hz

- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 100 μSEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
100 μW ≤ 5 μSEC; 0 TO 1 KHz
PW = 100 μSEC; 0 TO 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 - 60 Hz

[Signature]