


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5A-C-CA-EN-M-P*

S.N.: *5225*

Date: *NOV 16 '89*



$R_c = 0.75 \Omega$
PRF = 5 kHz

- a) Output signal amplitude:
0 TO + 40 AMPS
- b) Pulse width: *TO 0.75 ns*
20 TO 150 NSER
- c) Rise time:
 ≤ 5 NSER
- d) Fall time:
 ≤ 5 NSER
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
- g) Prime power:
O/K
120/240V
50-60 Hz