


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *5226*

Date: *NOV 15 1989*



40 dB ATTEN

10 VOLTS/DIV

PRF = 700 KHz

- a) Output signal amplitude:
≥ +30 VOLTS
- b) Pulse width:
≤ 1.0 nSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
150 mA (MAX)

AA