


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-A1-PW-C-N-451*

S.N.: *5223*

Date: *NOV 8 89*



*$R_L = 50 \Omega$
60 db ATTEN
 ≈ 100 VOLTS/DIV
PRT = 50 KHz*

- a) Output signal amplitude:
0 TO -100 V.
- b) Pulse width:
20 TO 50 NSOE
- c) Rise time:
53 NSOE
- d) Fall time:
65 NSOE
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*