



PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-3-pw-c-pn-4QA*
 S.N.: *5236*
 Date: *DEC 21 89*



50V
→


Pout RL = 100K



50V
→

Pout RL = 100K

50V
→



Nout RL = 100K

- a) Output signal amplitude: *0 TO ±150 V TO*
- b) Pulse width: *RL 7, 100K*
- c) Rise time: *0.1 us TO 1.0 us*
- d) Fall time: *≤ 20 nsec*
- e) PRF: *≤ 20 nsec*
- f) Jitter, stability: *0.1 Hz TO 1 kHz*
(10% MAX DATA CYCLE)
OK
- g) Prime power: *120/240 V*
50/60