


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANMN-3A-CT


S.N.: 5238

Date: NOV 17 89



- a) Output signal amplitude:
0 TO + 5 VOLTS
- b) Pulse width:
5 TO 40 NSER
- c) Rise time:
(3.2 μ MAX 174 M CYCLE)
- d) Fall time:
 ≤ 1 NSER
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
5.0-60 Hz

(A) 40 dB ATTEN
5 VOLTS/DIV
PRF = 12 MHz



(B) 40 dB ATTEN
5 VOLTS/DIV
PRF = 50 MHz

[Handwritten signature]