


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMR-1-C-N-ED-CA-ET-LA2*

S.N.: *5248*

Date: *OCT 26 1989*



R_L = 50 Ω
PRF = 50 KHz

- a) Output signal amplitude:
0 TO -5 VOLTS TO 50N
- b) Pulse width:
0.1 TO 0.5 μ SEC
- c) Rise time:
 \leq 1 NSEC
- d) Fall time:
 \leq 1 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz
- h) DELAY AD:
0 to \pm 1.0 μ SEC.