



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-A-C-PN*
 S.N.: *5251*
 Date: *DEC 19 89*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV
PRF ≈ 100 KHz



Ⓑ *AS Ⓐ BUT*
RISE TIME DETAIL

- a) Output signal amplitude:
0 TO ±15V
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz