


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APD-M-1-C-P*

S.N.: *5258*

Date: *JAN 12 1990*



40 db ATTEN

10 VOLTS DIV

PRF = 1.0 MHz

- a) Output signal amplitude:
0 TO 710 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]