


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-3-PW-PN-C-MONI*


S.N.: *5273*

Date: *DEC 15 1989*



- a) Output signal amplitude:
0 TO ±130 VOLTS
- b) Pulse width:
*TO R_L ≥ 100K
5.0 µS TO 0.5 SEC*
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
1 Hz TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60A*

Ⓐ *LOAD: ≈ OPEN CIRCUIT.
PRF NEAR MAX*



Ⓑ *LOAD: ≈ OPEN CIRCUIT.
PRF ≈ 20 Hz*