


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PNH-MV2-P-LIA*

S.N.: *5275*

Date: *NOV 7 1989*



50 db ATTEN

~ 32 VOLTS/DIV

PRF = 250 KHz

($R_L = 50 \Omega$)

a) Output signal amplitude:
0 TO + 80 VOLTS

b) Pulse width:
 ≤ 3 NSEC.

c) Rise time:
NA

d) Fall time:
NA

e) PRF:
0 TO 250 KHz

f) Jitter, stability:
OK

g) Prime power:
+15 \pm 1.0 VOLT.
200 mA (MAX)