


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-AV-1-C-P*


S.N.: *5278*

Date: *DEC 11 1989*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

*(A) 60 dB ATTEN
100 VOLTS/DIV*



*(B) 15 @ BUT
1 NS/DIV
(RISE TIME)*