


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2-PS-P*
S.N.: *5304*
Date: *NOV 24 1989*



- a) Output signal amplitude: *0 TO +10 V/AMP'S*
- b) Pulse width: *TO 5 μ S*
- c) Rise time: *2 TO 15 NSEC*
≤ 1 NSEC
- d) Fall time: *≤ 2 NSEC*
- e) PRF: *0 TO 20 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 HZ

120/240 V
50-60 HZ
[Signature]