


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVN-3-P
S.N.: 5305
Date: DEC 15 1989



- a) Output signal amplitude:
0 TO +15 VOLTS ($f \leq 100$ MHz)
- b) Pulse width:
0.2 TO 1.0 NSEC
- c) Rise time:
 ≤ 150 PSEC
- d) Fall time:
 ≤ 150 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 $+24 \pm 1.0$ VOLTS
300 mA (MAX)

40 dB ATTEN
10 VOLTS/DIV
PROF 110 MHz