


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M6-2-C-M-CA-N*

S.N.: *5311*

Date: *DEC 22 89*



50 db ATTN

2.32 VOLTS DIV

PRF = 50 KHz

- a) Output signal amplitude:
0 TO -125 V.
- b) Pulse width:
 $\le 2\text{ nsec}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz