


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AN-107B-P*  
S.N.: *5334*  
Date: *JAN 29 1990*



- a) Output signal amplitude:  
*0 TO +2 AMP*
- b) Pulse width:  
*TO 0 to 60 us*
- c) Rise time:  
*20 TO 200 nsec*
- d) Fall time:  
*≤ 10 nsec*
- e) PRF:  
*≤ 10 nsec*
- f) Jitter, stability:  
*0 TO 20 kHz*
- g) Prime power:  
*OK*

*R<sub>L</sub> = 25 Ω*  
*V = 20 VOLTS/DIV*  
*i = .4 AMPS/DIV*  
*PRF = 20 kHz*

*+24 VOLTS*  
*± 1 VOLT*  
*300 mA (MAX)*