


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-4V2-P-C EA*

S.N.: *5340*

Date: *FEB 20 90*



60 db ATTEN
100 VOLT/DIV
20 KHz (PRF)
 $R_c = 50 \Omega$

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
8 TO 20 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 \pm 1.0 VOLT
320 mA (MAX)

[Signature]