


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PS-05-CA-N*
S.N.: *5350*
Date: *FEB 20 1990*



R_L = 50 Ω
PRF = 1 kHz
50 VOLTS/DIV
0.5 USEC/DIV

- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width:
0.1 TO 100 USEC
- c) Rise time:
 ≤ 10 USEC
- d) Fall time:
 ≤ 10 USEC
- e) PRF:
M/A DW: 0 TO 50 kHz
PW ≤ 5 USEC: 0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 kHz

[Signature]