


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVA-HV1-C-P-05

S.N.: 5358

Date: FEB 14 90



50 db ATTEN
≈ 32 VOLTS/DIV
PRF ≈ 100 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz