


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVM-1-C-PFN-0T*

S.N.: *5361*

Date: *FEB 12 90*



Ⓐ *40 db ATTEN*
5 VOLTS/DIV
PRF ≈ 12 MHz



Ⓑ *40 db ATTEN*
5 VOLTS/DIV
PRF ≈ 25 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.1 TO 6.0 nSEC.
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 MHz

