


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PL-2A-OMI-P*
S.N.: *5373*
Date: *FEB 16 '90*



60 db atten
100 uV/B/DIU
PRF = 20 KHz

- a) Output signal amplitude:
0 to +150 V_{OUT}
- b) Pulse width:
5 to 100 nSEC
- c) Rise time:
52 nSEC
- d) Fall time:
53 nSEC
- e) PRF:
0 to 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
550 mA (MAX)