


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AV-143-C-P*

S.N.: *5377*

Date: *FEB 1 1990*



*50 dB ATTEN*

*32 VOLTS/DIV*

*PRF = 1.0 MHz*

a) Output signal amplitude:

*0 TO +40 VOLTS*

b) Pulse width:

*0.4 TO 2.0 NS*

c) Rise time:

*≤ 150 PSEZ*

d) Fall time:

*≤ 250 PSEZ*

e) PRF:

*0 TO 1.0 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240 V*

*50-60 Hz*