


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVH-HV3-P-05A*

S.N.: *5379*

Date: *MAR 9 1990*



*50 db ATTEN*  
*32 VOLT/DIV*  
*PRF = 50 KHz*

a) Output signal amplitude:  
*≥ +60 VOLTS*

b) Pulse width:  
*≤ 1 NSEC*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 1 MHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+15 ± 1.0 VOLT*  
*100 mA (MAX)*