


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV-107E-PS-IBM4*

S.N.: *5382*

Date: *FEB 23 90*



- a) Output signal amplitude: *0 TO +0.5 A TO*
- b) Pulse width: *50  $\mu$ s*  
*100 NSR TO 1.0  $\mu$ s*
- c) Rise time:
- d) Fall time:
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240V*  
*50 to 60 Hz*

*R<sub>L</sub> = 50  $\Omega$*   
*10 VOLTS/DIV*  
*200 mA/DIV*  
*200 NSR/DIV*  
*PRF = 10 KHz*