


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AN-2-C-CA-EN-PN-M*
S.N.: *5386*
Date: *MAR 20 90*



- a) Output signal amplitude:
0 TO ±350 V
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

70 db ATTEN
≈ 320 VOLTS/Div
PRF = 2.0 KHZ

120/240 V
50-60 Hz
[Signature]