


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MM-2-C-PDN-D*

S.N.: *5393*

Date: *MAR 22 '90*



- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 150 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz

40 db ATTN

10 VOLTS DIV

PRF ≈ 12 MHz