


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-3-PW-PN-C-TKEFA

S.N.: 5401

Date: APRIL 24 1990



a) Output signal amplitude:

0 TO  $\pm 220$  VOLTS

b) Pulse width:

100 USEC TO 10 MSEC \*

c) Rise time:

$\leq 100$  NSEC

d) Fall time:

$\leq 100$  NSEC

e) PRF:

\*\*

50 HZ TO 5 KHZ

f) Jitter, stability:

OK

g) Prime power:

170/240 V, 50-60 HZ

\* USE-FUL TO 100 MSEC

\*\* USE-FUL TO 5 HZ

a 50 VOLTS/DIV

200 USEC/DIV

$R_L = 100$  K.