


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PW-C-N-05*


S.N.: *5403*

Date: *MAR 14 90*



- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width: *TO 50 n*
0.1 TO 100 USEC.
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW: 0 TO 50 Hz
PW ≤ 5 USEC: 0 TO
- f) Jitter, stability: *1 kHz*
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

Ⓐ R_L = 50 Ω
50 VOLTS/DIV
20 USEC/DIV
PRF = 50 kHz



Ⓑ
50 VOLTS/DIV
200 NSEC/DIV
PRF = 1 kHz