


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-C-M-P*
S.N.: *5406*
Date: *MAR 5 90*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0.2 TO 2.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz
[Signature]