


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-1*

S.N.: *5412*

Date: *MAR 12 90*



- a) Output signal amplitude:
+25 to +100 mA
- b) Pulse width:
≤ 25 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 kHz
- f) Jitter, stability:

OK

*≈ 60 amps/div
PRF = 500 Hz*

- g) Prime power:
*+15 ± 1.0 volt
250 mA (max)*