


PULSE GENERATOR  
PERFORMANCE CHECK

Model: A20-1  
S.N.: 5414  
Date: MAR 12 '90



- a) Output signal amplitude:  
+25 TO +100 AMP
- b) Pulse width:  
≤ 25 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 VOLT  
250 mA (MAX)

≈ 60 AMPs PK  
PRF = 500 Hz

*[Signature]*