


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-C-P-TAMB*

S.N.: *5415*

Date: *APRIL 11 1990*



*$R_L = 10K$
1.0 μ SEC/DIV
(≈ 250 KHz)
50 VOLTS/DIV*

- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: *TO $R_L > 10K$,
2 TO 20 μ SEC*
- c) Rise time: *C50% DUTY
 ≤ 5 NSEC*
- d) Fall time:
 ≤ 5 NSEC
- e) PRF: *25 KHz TO
250 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50-60 Hz*