


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-A-1-pw-CP*

S.N.: *5420*

Date: *MAR 29 80*



- a) Output signal amplitude:
0 TO + 200 VOLTS
- b) Pulse width:
50 NS
- c) Rise time: *1.0 USEC. (0.5%
MAX DATA CYCLE)*
- d) Fall time:
≤ 10 NS
- e) PRF:
MIN PW: 0 TO 100 KHz
MAX PW: 0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

V_L = 50 m
PRF = 5 KHz

120/240 V
50-60 Hz
[Signature]