


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-A-2-PW-C-PN-05

S.N.: 5422

Date: APRIL 17 1990



- a) Output signal amplitude:  
0 TO  $\pm 50$  VOLTS
- b) Pulse width:  
20 TO 200 NS62
- c) Rise time:  
 $\leq 5$  NS62
- d) Fall time:  
 $\leq 5$  NS62
- e) PRF: 0 TO 10 MHz  
(20% MAX DUTY CYCLE)
- f) Jitter, stability:  
OK
- g) Prime power:

$R_i = 50 \Omega$

100 NS62/DIV

10 VOLTS/DIV

PRF  $\approx$  100 KHz

120/240 V

50-60 Hz