


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5D-C-PN-M*


S.N.: *5424*

Date: *APRIL 5 1990*



- a) Output signal amplitude:  
*0 TO ±30 AMP*
- b) Pulse width: *TO 3 μs*
- c) Rise time:  
*0.1 TO 5.0 μs*  
*≤ 40 NS*
- d) Fall time:  
*≤ 40 NS*
- e) PRF:  
*0 TO 300 Hz*
- f) Jitter, stability:  
*OK*

*Point* *0.5 μs/div*  
*50 V/div*  
*R<sub>L</sub> = 3.0 Ω*  
*PRF = 300 Hz*



- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*Point* *0.5 μs/div*  
*50 V/div*  
*R<sub>L</sub> = 3.0 Ω*  
*PRF = 300 Hz*

*[Signature]*