

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-1-PS-P-OT-M*

S.N.: *5430*

Date: *APRIL 26 1990*



120 db ATTEN
5 VOLTS/DIV

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz