


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PL-2A-C-CM1-P*

S.N.: *5439*

Date: *JUNE 27 1990*



60 db ATTEN
.100 VOLTS/DIV
PRF = 20 KHz

- a) Output signal amplitude:
0 TO +170 VOLTS
- b) Pulse width:
0 TO +150 VOLTS AT 20 KHz
5 TO 100 NSEC
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]