


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVC-C-PN-SLACI*

S.N.: *5449*

Date: *MAY 18 1990*



*(A) Pout 50 db atten
~ 32 VOLTS/DIV*



*(B) AS (A) BUT
N out*

- a) Output signal amplitude:
≥ 120 V (PP)
- b) Pulse width:
T/2 = 0.8 nSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V.
50-60 Hz*