


PULSE GENERATOR  
PERFORMANCE CHECK


Model: AVI-U-HV3-C-OT-PN-CA-EN

S.N.: 5451

Date: MAR 10 1990



(A) 60 db ATTEN  
100 VOLTS/DIV  
(RISE TIME)



(B) AS (A) BUT  
10 NSER/DIV  
(PRF ≈ 5.0 KHz)

- a) Output signal amplitude:  
0 TO ±100 VOLTS
- b) Pulse width:  
4 TO 100 NSER
- c) Rise time:  
≤ 200 PSEZ
- d) Fall time:  
≤ 4 NSER
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

120 / 240 V  
50 - 60 Hz