



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-B-P-M-0A-EW-00*
 S.N.: *5457*
 Date: *MAY 26 1990*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NSER
- c) Rise time:
≤ 100 PSEK
- d) Fall time:
≤ 135 PSEK
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz

Ⓐ *40 db ATTEN*
10 VOLTS/DIV
PRF ≈ 7 MHz



Ⓑ *40 db ATTEN*
10 VOLTS/DIV
PRF ≈ 25 MHz

[Signature]