


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-PS-7-M-CA EW-60*  
S.N.: *5458*  
Date: *MAY 26 1990*



- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*2 TO 15 NS*
- c) Rise time:  
*≤ 150 PSEZ*
- d) Fall time:  
*≤ 600 PSEZ*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*PRF ≈ 7 MHz*

- g) Prime power:  
*120/240 V.*  
*50-60 MHz*

*[Signature]*