


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVN-3-C-OT-PN-M*

S.N.: *5463*

Date: *MAY 9 1990*



*40 dB ATTEN
10 VOLTS/DIV
PRF \approx 100 MHz*



*40 dB ATTEN
10 VOLTS/DIV
PRF \approx 200 MHz*

- a) Output signal amplitude:
0 TO ± 15 VOLTS ($f \leq 100$ MHz)
- b) Pulse width:
*0 TO ± 5 VOLTS ($f \leq 250$ MHz)
0.2 TO 1.0 NS*
- c) Rise time:
 ≤ 150 PSEC
- d) Fall time:
 ≤ 150 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*