


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-PS-P*
S.N.: *5472*
Date: *APRIL 16 1990*



70 db ATTEN
320 VOLTS/DIV
PRF = 2.0 KHz

- a) Output signal amplitude:
0 TO +350. 100mV
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 2 nSEC
- d) Fall time:
≤ 3 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz