


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-AV-1-C-PN

S.N.: 5476

Date: APRIL 23 1990



60 db ATTEN
100 VOLT/DIV

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
3 TO 100 NS
- c) Rise time:
 ≤ 1 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ

