


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-P*  
S.N.: *5485*  
Date: *MAY 25 1990*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NS*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*  
*5 VOLTS/DIV*  
*PRF ≈ 7 MHz*

*+24 ± 1.0 VOLT*  
*400 mA*  
*MAX*  
*[Signature]*