


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVNN-1-PS-P-OT

S.N.: 5501

Date: AUG 7 1990



a) Output signal amplitude:

0 TO 5 VOLTS

b) Pulse width:

1 TO 30 NS

c) Rise time: (30 TO 70% DUTY CYCLE LIMIT)

≤ 100 PSEC

d) Fall time:

≤ 100 PSEC

e) PRF:

25 TO 250 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 HZ

40db ATTEN.

5 VOLTS/DIV

V/F \approx 100 MHz