


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-134-C-P-M2*

S.N.: *5503*

Date: *JUN 7 30 1990*



60 66 ATTEN
100 VOLTS/DIV
PRF = 1.0 MHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
10 TO 20 NS
- c) Rise time:
61 NS
- d) Fall time:
63 NS
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz