


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-PS-PN-05

S.N.: 5517

Date: JULY 12 1990



70 dB ATTEN
~ 320 VOLTS/DIV
PRF ~ 2 KHz

- a) Output signal amplitude:
0 TO ± 350 VOLTS
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
 ≤ 2 NSER
- d) Fall time:
 ≤ 3 NSER
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.