


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *A20-3B-F*

S.N.: *5518*

Date: *JUNE 8 1990*



*60 db ATTEN*

*100 VOLTS/DIV*

*$R_L \approx 50 \Omega$*

*$\approx 20 \text{ AMPS/DIV}$*

a) Output signal amplitude:

*+ 10 TO +30 AMPS*

b) Pulse width:

*1 ± 0.5 NS*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 20 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+ 15 VOLTS*

*200 mA (LIMIT)*

*[Signature]*